

ALC-0660-05000-FM400.22 SETIA

Overview

Main Features

- 660 nm / 5.0 W
- 400 μm core fiber
- Detachable fiber
- Compact footprint
- Fiber sensor (optional)
- Temperature sensor
- Power monitor (optional) •
- Visible pointer (optional) •

Typical Performance Characteristics					
Parameter	Symbol	LD	Unit		
Operating temperature	Т	20	°C		
Operating power ex-fiber	Pop	5.0	W		
Center wavelength	λ	660±8	nm		
Spectrum width, 1/e ²	Δλ	<6	nm		
Wavelength-temperature coefficient	Δλ/ΔΤ	0.17	nm/ºC		
Linearized wavelength-current	Δλ/ΔΙ		nm/A		
Threshold current	I _{th}	0.65	А		
Differential efficiency	η _{ext}	5.3	W/A		
Operating current	I _{op}	1.65	А		
Operating voltage	V _{op}	12.1	V		
Maximum current (10 seconds)	I _{max}	1.8	А		
Detachable fiber connector		SMA			
Fiber core diameter	D _{fiber}	400	μm		
Fiber NA		0.22			

OPTIONAL POINTER: >1mW at 638±5nm, 2.9V operating voltage, typical drive current <30mA





V-I at 20 °C

14

12

10

8

6

4

2

0 0.0

0.5

1.0

1.5


Current, A

2.0

2.5


Voltage, V

Electro-optical characteristics





Dimensions and Pinout




Mechanical Drawings

AKELA Laser Corporation, 1095 Cranbury South River Road, Suite 14, New Jersey 08831, USA Phone: (732) 305-7105; email: info@AKELALaser.com; web: www.AKELALaser.com



Pinout



Pin #	Function
1	Red pointer (+)
2	Red pointer (-)
3-7	Not used
8	Laser diode (+)
9-10	Not used
11 & 12	Thermistor outputs
13	Power monitor photodiode (+)
14-18	Not used
19	Laser diode (-)
20	Power monitor photodiode (-)
21	Fiber Sensor Phototransistor Collector
22	Fiber Sensor Phototransistor Emitter
23	Fiber Sensor LED Cathode
24	Fiber Sensor LED Anode

AKELA Laser Corporation, 1095 Cranbury South River Road, Suite 14, New Jersey 08831, USA Phone: (732) 305-7105; email: info@AKELALaser.com; web: www.AKELALaser.com



Document Revision History

Revision Number	Revision Date	Nature of Revision	Approved by
1	April 8, 2024	New document created	М.М.



These components do not comply with the Federal Regulations (21 CFR Subchapter 1) as administered by the Center for Devices and Radiological health. Purchaser acknowledges that his/her products must comply with these regulations before they can be sold. Akela laser Corporation reserves right to change any specifications.